

# Test Report 3219386.


## Jiangxi Rongge Technology Co., Ltd.

## Introduction.

This report has been prepared by Paul Waller and relates to the activity detailed below:

	Client Details
<b>Job number: 3219386</b>	Jiangxi Rongge Technology Co., Ltd.
Job type: Testing Samples Submitted	1# Xiaofeng Road, Xiaofeng Industrial Park
Start Date: 15/05/2020	Economic Development Zone
Test type: Type retest	Guangfeng District
Sample ID: 10190140	Shangrao City
<b>Registration: CE 728056</b>	Jiangxi
	334604

The report has been approved for issue by T Wicksey – Senior Test Engineer

Approved For Issue	
	Issue Date: 21 May 2020

## Objectives.

This is an independent test evaluation to only certain clauses or sub-clauses of the agreed specification in accordance with the following test programme:

BSI COVID-19 filtering face piece technical specification, for COVID-19 masks for use by healthcare workers

## Product Scope.

COVID-19 masks for use by healthcare workers

## Report Summary.

The samples were received on 7 May 2020 and the testing was started on 15 May 2020.

The samples submitted complied with the requirements of the test work conducted.

## Test Samples.

Sample ID	ER Number	Description
1 to 19	10190140	Model: RG-MX-P01 FFP2

## Description of Test Samples.

Sample Description
COVID-19 masks for use by healthcare workers: Model: RG-MX-P01 FFP2

# Test Requirements.

## Testing in accordance with BSI COVID-19 filtering face piece technical specification

Technical testing specification for COVID-19 masks for use by healthcare workers

EN 149:2001+A1:2009 Performance requirement	EN 149:2001+A1:2009 Test method clause	Requirement	Assessment
<b>7.7 Practical performance</b> The particle filtering half mask shall undergo practical performance tests under realistic conditions. These general tests serve the purpose of checking the equipment for imperfections that cannot be determined by the tests described elsewhere in this standard. Where practical performance tests show the apparatus has imperfections related to wearer's acceptance, the test house shall provide full details of those parts of the practical performance tests which revealed these imperfections.  2 test subjects, masks tested 'As received'	Testing shall be done in accordance with 8.4.	During the tests the particle filtering half mask shall be subjectively assessed by the wearer and after the test, comments on the following shall be recorded: a) head harness comfort; b) security of fastenings; c) field of vision; d) any other comments reported by the wearer on request.	Pass
<b>7.9 Leakage</b> <b>7.9.1 Total inward leakage</b> 5 test subjects, masks tested 'As Received'	Testing shall be done in accordance with 8.5.	All samples must achieve All individual exercise results tests shall be not greater than 11 % (for FFP2) and, in addition, all arithmetic means for the total inward leakage shall be not greater than 8 % (for FFP2)	Pass
<b>7.9 Leakage</b> <b>7.9.2 Penetration of filter material</b> 3 test samples masks tested 'As Received' for NaCl (Sodium Chloride) and PO (Paraffin oil), 3min test	Testing shall be done in accordance with 8.11	6% for both PO and NaCl	Pass
<b>7.12 Carbon dioxide content of the inhalation air</b> 3 test samples, masks tested 'As Received'	Testing shall be done in accordance with 8.7.	The carbon dioxide content of the inhalation air (dead space) shall not exceed an average of 1,0 % (by volume).	Pass
<b>7.16 Breathing resistance</b> 3 test samples, masks tested 'As Received'	Testing shall be done in accordance with 8.9	The breathing resistances shall meet the requirements of; 30l/min - 0.7mbar (inhale) 95l/min - 2.4mbar (inhale) 160l/min - 3.0mbar (exhale)	Pass
<b>Appendix A – Test Panel Data</b>			
<b>Product Photographs</b>			

## Glossary of Terms.

Pass: Complies. Tested by BSI engineers at BSI laboratories

Pass 1: Complies. Witnessed by BSI engineers in manufacturers laboratory.

Pass 2: Complies. Tests carried out by third party lab; results accepted by BSI.

Pass\*: Report resulted in uncertainty and states that Compliance is more probable than non-compliance.

Fail: Non-compliance. Product does not meet the requirements of this clause.

Fail\*: Report resulted in uncertainty and states that Non-compliance is more probable than compliance.

N/T: Not Tested

N/A: Not Applicable

AR: As Received

TC: Temperature Conditioned

SW: Simulated Wear

FT: Flow Tested

MS: Mechanical strength

MMDF: Manufactures Minimum Design Flow

MMDC: Manufactures Minimum Design Condition

## Conditions of Issue.

This Test Report is issued subject to the conditions stated in current issue of 'BSI Terms of Service'. The results contained herein apply only to the particular sample(s) tested and to the specific tests carried out, as detailed in this Test Report. The issuing of this Test Report does not indicate any measure of Approval, Certification, Supervision, Control or Surveillance by BSI of any product.

No extract, abridgment or abstraction from a Test Report may be published or used to advertise a product without the written consent of BSI, who reserve the absolute right to agree or reject all or any of the details of any items or publicity for which consent may be sought.

Should you wish to speak with BSI in relation to this report, please contact Customer Services on +44 (0)8450 80 9000.

BSI

Kitemark House  
Maylands Avenue  
Hemel Hempstead  
Hertfordshire  
HP2 4SQ



0135

Opinions and Interpretations expressed herein are outside the scope of our UKAS accreditation.

Unless otherwise stated, any results not obtained from testing in a BSI laboratory are outside the scope of our UKAS accreditation.

# Test Results.

## Testing in accordance with BSI COVID-19 filtering face piece technical specification

BS EN 149:2001 +A1:2009 Technical testing specification for COVID-19 masks for use by healthcare workers

CLAUSE	REQUIREMENTS	ASSESSMENT
7.7	<b>Practical performance</b> The particle filtering half mask shall undergo practical performance tests under realistic conditions. These general tests serve the purpose of checking the equipment for imperfections that cannot be determined by the tests described elsewhere in this standard. Where practical performance tests show the apparatus has imperfections related to wearers acceptance, the test house shall provide full details of those parts of the practical performance tests which revealed these imperfections. Test in accordance with clause 8.4 of the standard.	Pass

### Testing in accordance with BSI COVID-19 filtering face piece technical specification, for masks for use by healthcare workers

During the tests the particle filtering half mask shall be subjectively assessed by the wearer and after the test, comments on the following shall be recorded:

a) head harness comfort; b) security of fastenings; c) field of vision; d) any other comments reported by the wearer on request.

**Table A:** Practical performance

Test candidate	Sample	Comments				Assessment
		Head harness comfort	Security of fastenings	Field of vision	Any other comments	
JS3	1 AR	OK	OK	Good	None	Pass
LM2	2 AR	OK	OK	OK	None	Pass

### Leakage

7.9.1

Total inward leakage

The laboratory tests shall indicate that the particle filtering half mask can be used by the wearer to protect with high probability against the potential hazard to be expected.

The total inward leakage consists of three components: face seal leakage, exhalation valve leakage (if exhalation valve fitted) and filter penetration.

Test in accordance with clause 8.5 of the standard.

### Testing in accordance with BSI COVID-19 filtering face piece technical specification, for masks for use by healthcare workers

5 test subjects, masks tested As received. All individual exercise results tests shall be not greater than 11 % (for FFP2) and, in addition, all arithmetic means for the total inward leakage shall be not greater than 8 % (for FFP2).

**Table B:** Clause 7.9.1 - Total inward leakage

Test candidate	Sample	Pre test condition	Inward Leakage (%)						Assessment
			A	B	C	D	E	Average	
			Walking	Walking with head side to side	Walking with head up &down	Walking and talking	Walking		
BH1	3	AR	5.58	6.07	8.42	3.95	6.06	6.02	Pass
JB1	4	AR	1.16	1.48	1.92	2.46	2.80	1.96	Pass
RF1	5	AR	1.22	9.96	9.85	2.55	7.14	6.14	Pass
GR1	6	AR	4.38	4.22	4.35	1.24	4.04	3.65	Pass
BH2	7	AR	2.31	2.06	2.01	0.38	0.45	1.44	Pass

# Test Results. (Continued)

CLAUSE	REQUIREMENTS	ASSESSMENT
--------	--------------	------------

7.9.2 Penetration of filter material

**Testing in accordance with BSI COVID-19 filtering face piece technical specification, for masks for use by healthcare workers**

3 test samples masks tested As received; for NaCl (Sodium Chloride) and PO (Paraffin oil), 3 min test. Testing shall be done in accordance with 8.11. 6% limit for both PO and NaCl

Pass

**Table C:** Clause 8.11 - Sodium Chloride penetration test

Sample number	Pre-test condition	Flow through filter (l/min)	Penetration (%)	
			Limit	Actual
8	AR	95	< 6	0.278
9	AR			0.315
10	AR			0.285

**Table D:** Clause 8.11 - Paraffin oil penetration test

Sample number	Pre-test condition	Flow through filter (l/min)	Penetration (%)	
			Limit	Actual
11	AR	95	< 6	2.258
12	AR			2.259
13	AR			0.848

7.12

**Carbon dioxide content of inhalation air**

The carbon dioxide content of the inhalation air (dead space) shall not exceed an average of 1.0% (by volume).

Pass

Test in accordance with clause 8.7 of the standard.

**Table E:** Clause 8.7 - Carbon Dioxide content of the inhalation air

Sample	Pre-test condition	Dead space CO <sub>2</sub> (%)	
		Limit	Measured
14	AR	< 1.0	0.57
15	AR		0.60
16	AR		0.60

# Test Results. (Continued)

CLAUSE	REQUIREMENTS	ASSESSMENT
--------	--------------	------------

7.16

## Breathing resistance

Testing in accordance with BSI COVID-19 filtering face piece technical specification, for masks for use by healthcare workers

3 test samples masks tested 'As received'. Test in accordance with clause 8.9 of the standard.

Pass

The breathing resistances shall meet the requirements of FFP2;

30l/min - 0.7mbar (inhale), 95l/min - 2.4mbar (inhale), 160l/min - 3.0mbar (exhale)

**Table F:** Clause 8.9 - Breathing resistance. Inhalation resistance at a continuous flow

Sample	Pre-test condition	Continuous flow (l/min)	Inhalation resistance (mbar)	
			Limit	Measured
17	AR	30	< 0.7	0.44
18	AR			0.47
19	AR			0.45
17	AR	95	< 2.4	1.38
18	AR			1.44
19	AR			1.40

**Table G:** Clause 8.9 - Breathing resistance. Exhalation resistance at a continuous flow, measured in five orientations with the worst case reported

Sample	Pre-test condition	Continuous flow (l/min)	Exhalation resistance (mbar)	
			Limit	Measured
17	AR	160	< 3.0	2.16
18	AR			2.26
19	AR			2.15



## Appendix A. – Test Panel Data

Test Candidate	Facial Dimensions (mm)					Sex
	Length of face	Width of face	Face depth	Width of mouth	Head Circumference	
JS3	126	134	124	49	600	Male
LM2	110	148	125	44	589	Male
BH1	124	145	126	49	590	Male
JB1	114	144	108	59	574	Male
RF1	104	122	121	55	549	Male
GR1	124	145	126	49	590	Male
BH2	124	148	120	51	595	Male

Note: All candidates were clean shaven

## Product photographs.



Front view



Side View



Inside View

\*\*\*End of Report\*\*\*